

World Headquarters



Millar, Inc.
6001-A Gulf Freeway
Houston, Texas 77023 USA
Phone: 832-667-7000 or 800-669-2343 (in the USA)
Fax: 713-714-8497
Email: info@millar.com
Web site: millar.com

Millar Worldwide Distribution

Millar, Inc. has a network of Authorized Distributors in most countries around the world. For information on the Millar distributor in your country, please contact the Millar Customer Service Department at our headquarters in Houston, TX.



Making the
improbable possible.

**Interface Cable
Models AEC-10C & AEC-10D
For
ADInstruments Bridge Amplifier**

Animal Use Only

Instructions for Use

© 2007, 2013 Millar, Inc. All rights reserved
Millar and Mikro-Tip are registered trademarks of Millar, Inc.
Products and company names used are the trademarks or trade names of their respective companies.

M.I. P/N: 004-2140 Rev. D

Millar Limited Warranty

Millar, Inc. warrants all of its cables to be free from defects in workmanship and materials at the time of shipment to the original purchaser.

Millar hereby excludes all warranties not herein stated, whether express or implied by operation of law or course of dealing or trade usage or otherwise, including but not limited to any implied warranties of fitness or merchantability.

Since handling, storage, cleaning and sterilization of the product, as well as factors relating to patient diagnosis, treatment, catheterization procedures, and other matters beyond Millar's control, directly affect the product and the results obtained from its use, Millar shall not be liable for any incidental or consequential loss, damage, or expense arising directly or indirectly from the use of this product.






Specifications subject to change without notice.

READ INSTRUCTIONS FIRST

Device Description

The AEC-10C and AEC-10D cables are designed to be an interface between Millar pressure transducer catheters and the ADInstruments bridge amplifier. The AEC-10C has a viking connector and the AEC-10D has a low profile connector. These cables are for Animal Use Only.

This product is designed for use by professionals with appropriate education and training in life science and medical research applications.

Definition of Symbols	
	Attention, consult accompanying documents
	Date of Manufacture
REF	Catalog Number
SN	Serial Number
	Batch Code
	EU Declaration of Conformity
	Waste Electrical and Electronic Equipment

Warnings

- Customer is responsible for EMC compliance when this cable is attached to their input device.

Cleaning

Wipe the cable and connectors clean with a soft wet gauze or tissue. If extremely dirty, the cable may be soaked in a solution of enzymatic cleaning solution. **DO NOT SOAK OR CLEAN WITH HARSH CHEMICALS SUCH AS ACETONE!**

CAUTION: Do not immerse the electrical connectors.

Recommended Sterilization Method (Optional)

In most cases sterilization is not required and should only be used in the event of a possibility of infection between breeds of rodents due to microbial forms such as bacterial spores on the device. If an infection is possible, the use of a high level disinfectant is recommended since this process is less lethal and destroys most recognized pathogenic microorganisms. The use of Cidex OPA, Cidex Activated or Metricide disinfectant is recommended in lieu of ETO sterilization. If protocol mandates sterilization, the following cycle parameters are to be used.

Ethylene Oxide Gas: The ethylene oxide sterilizer should be thoroughly cleaned before each sterilization cycle. The cables should be completely dry before sterilization as water on the units may react with ethylene oxide and reduce its effectiveness.

Sterilization Cycle Parameters

Preheat phase: Starting Temperature 110 °F (43 °C) min.
Duration 30 minutes

Initial Vacuum: 6.0 inHgA (20.3 kPa)
Rate: 3 minutes

Nitrogen Flush: 2 cycles
Nitrogen Addition to: 28.0 ± 0.5 inHgA (94.8 ± 1.7 kPa)
Rate: 1.4 ± 0.5 inHgA/min. (4.7 ± 1.7 kPa/min.)
Evacuation: 6.0 ± 0.5 inHgA (20.3 ± 1.7 kPa)
Rate: 1.0 ± 0.5 inHgA/min. (3.4 ± 1.7 kPa)

Conditioning

Humidification: 1.5 ± 0.5 inHgA (5.1 ± 1.7 kPa)
Steam Conditioning: 10 min. + 5 min. - 2 min.
Humidity Dwell: 30 ± 5 min.
Relative Humidity: 7.5 ± 0.5 inHgA (25.4 ± 1.7 kPa)
15-70%

Ethylene Oxide Concentration: 500 ± 50 mg/L
Dwell Pressure: 16.5 ± 1.0 inHgA (55.8 ± 3.4 kPa)
Dwell Time: 2 hours - 1.0 + 30 min.
Temperature: 110-130 °F (43-54 °C)
Relative Humidity: 30-70%
(35-44% nominal)

After Vacuum

Vacuum: 6.0 ± 0.5 inHgA (20.3 ± 1.7 kPa)
Rate: 1.0 ± 0.5 inHgA/min. (3.4 ± 1.7 kPa)
Vacuum Hold: 10 min.

Gas Wash A:

Release: 28.0 inHgA/min. (94.8 ± 1.7 kPa)
Rate: 1.4 ± 0.5 inHgA/min. (4.7 ± 1.7 kPa)
Vacuum: 6.0 ± 0.5 inHgA (20.3 ± 1.7 kPa)
Rate: 1.0 ± 0.5 inHgA/min. (3.4 ± 1.7 kPa)

Release (Filtered Air) : 28.0 ± 0.5 inHgA (94.8 ± 1.7 kPa)
Rate: 2.0 ± 0.5 inHgA/min. (6.6 ± 1.7 kPa)

Aeration

Duration: At least 8 hours
Temperature: 110 ± 10 °F (43 °C)

Sterilization by autoclaving, ionizing (gamma) radiation, plasma, hydrogen peroxide, and exposure to formaldehyde vapor solutions are known to be incompatible with the cables and must not be used. Use of a Glutaraldehyde Solution (e.g., Cidex®) is not recommended because the electrical connectors cannot be immersed and thus would not be sterilized by this method.

Routine Inspection

Conductive liquid entering the cable connectors can cause erratic operation and possible corrosion of wires within the connectors. The connector pins should be routinely inspected for corrosion.

Schematics

