## Pulmonary Artery Pressure Data Comparison between the Millar Mikro-Cath™ and a Fluid-Filled Transducer

Pulmonary Artery Pressure (PAP) is one of the most commonly measured parameters during a cardiac catheterization case. Mean PAP, systolic PAP and diastolic PAP are often derived by visually marking the waveform output by a fluid-filled transducer. This waveform is influenced by a number of factors including patient position, transducer level, air bubbles, open connections and catheter tubing length; all of which can introduce error into the measurement. The result is a potential for over-estimation or under-estimation of the systolic and diastolic values. The Millar Mikro-Cath disposable pressure catheter can be introduced into the pulmonary artery via a 6F or larger multi-purpose catheter to deliver a more accurate and reproducible PAP waveform, independent of the factors which impact fluid-filled readings. The result is a simplification in both the recording and reading of PAP derived parameters.

The waveforms below highlight the differences between the Mikro-Cath and fluid-filled pressure transducers. The recordings were made in the same patient several minutes apart and have been overlaid for comparison purposes.



Millar Sensors.Systems.Solutions.@ Millar Instruments, Inc. | 6001-A Gulf Freeway | Houston, TX 77023 USA Ph. +1-832-667-7000 | 1-800-669-2343 (in USA) sales@millar.com | Millar.com

© 2012 Millar Instruments, Inc. All rights reserved. Millar, Mikro-Tip and Sensors. Systems. Solutions. are registered trademarks of Millar Instruments. Mikro-Cath is a trademark of Millar Instruments.